

Contract Number: 800519

Certificate of Compliance: SGSNA/19/SUW/00111
Certificate Project Number: SUW-CERT190600282



Certified Product: patient handling devices
Trademarks (if any): BestLift; BestStand
Model(s): PL182, PL228, SA182, SA182H, SA228, SA228H, PL400, PL500, PL600, SA400, SA500, SA600, PL400H, PL400HE, SA400HE, PL400EF, PL350CT, SA400H
Technical Data: Input 100-240V~, 50/ 60; Hz, 1A Output 28VDC 1.5A IP54 rated control box and actuator

Name and Address of Certificate Holder: Bestcare LLC
2920 Pacific Drive, 30071, Norcross, Georgia, United States of America

Name and Address of Manufacturer: Bestcare LLC
2920 Pacific Drive, 30071, Norcross, Georgia, United States of America

This certificate supercedes previous certificates issued with the same certificate number. Certification is valid when products are indicated on the SGS directory of certified products at www.sgs.com. The product is certified according to ISO/IEC Guide 17067, *Conformity assessment - Fundamentals of product certification, System 3*, and in accordance with

ANSI/AAMI ES60601-1:2005 +C1:2009 +A2:2010
CAN/CSA C22.2 No. 60601-1:14
ISO 10535:2006
ANSI/AAMI HA60601-1-11:2015
CAN/CSA C22.2 No. 60601-1-11:15

Authorized by Effective date: 09 July 2019


Paul Krauss
Certifier

Consumer and Retail Services, a division of SGS North America Inc.
620 Old Peachtree Road, Ste. 100, Suwanee, GA 30024, USA
t +1 770 570 1800 f +1 770 277 1240 www.sgs.com
Page 1 of 3

